

FujiFilm NH3 Quality control

For the control for clinical chemistry parameters of Element DC
FOR VETERINARY USE ONLY!



Item code 1528
Lot 20380
Expiry date 10/25
Handling: ready to use

Element CM XT / Fuji series	Parameter	Target value	Target range	conv. unit	Target value	Target range	SI unit
	NH3	56	40 - 73	μmol/L	79	56-102	μg/dl

