

AxonControl N/P

REF 77708/6 6 x 2 ml

Spotchem 4430



118702

2020-09

Version 6

Bestandteil Constituant Composante Constituent	Einheit Unité Unita Unit	AxonControl N (77706 Lot. 1135UNCM)			AxonControl P (77707 Lot. 879UECM)		
		Sollwert Valeur cible Valore nominale Assay value	Range (+/- 3s) Qualab ¹	Range (+/- 3s) Rililab ²	Sollwert Valeur cible Valore nominale Assay value	Range (+/- 3s) Qualab ¹	Range (+/- 3s) Rililab ²
Enzymes							
ALP	U/l	170	139 - 201	151 - 189	618	507 - 729	547 - 689
	µkat/l	2.83	2.32 - 3.34	2.52 - 3.15	10.3	8.4 - 12.2	9.1 - 11.5
ALT / GPT	U/l	23.0	17.0 - 29.0	20.4 - 25.6	58.0	47.6 - 68.4	51.3 - 64.7
	µkat/l	0.38	0.28 - 0.48	0.34 - 0.43	0.97	0.79 - 1.14	0.86 - 1.08
AST / GOT	U/l	29.5	23.5 - 35.5	26.1 - 32.9	66.0	54.1 - 77.9	58.4 - 73.6
	µkat/l	0.49	0.39 - 0.59	0.44 - 0.55	1.10	0.90 - 1.30	0.97 - 1.23
AMY	U/l	168	138 - 198		320	262 - 378	
	µkat/l	2.80	2.30 - 3.30		5.33	4.37 - 6.29	
CK / CPK	U/l	160	131 - 189	142 - 178	352	289 - 415	313 - 391
	µkat/l	2.67	2.19 - 3.15	2.37 - 2.96	5.87	4.81 - 6.92	5.22 - 6.51
GGT	U/l	30.0	22.0 - 38.0	26.6 - 33.5	125	102.5 - 148	110.6 - 139
	µkat/l	0.50	0.37 - 0.63	0.44 - 0.56	2.08	1.71 - 2.46	1.84 - 2.32
LDH / LD IFCC	U/l	321	263 - 379	292 - 350	502	412 - 592	457 - 547
	µkat/l	5.35	4.39 - 6.31	4.87 - 5.83	8.37	6.86 - 9.87	7.61 - 9.12
Substrates							
ALB	g/l	33.5	29.5 - 37.5	29.3 - 37.7	46.0	40.5 - 51.5	40.3 - 51.8
	g/dl	3.35	2.95 - 3.75	2.93 - 3.77	4.60	4.05 - 5.15	4.03 - 5.18
T-BIL / BIL	µmol/l	16.0	13.1 - 18.9	12.5 - 19.5	66.0	54.1 - 77.9	57.4 - 74.6
	mg/dl	0.94	0.77 - 1.10	0.73 - 1.14	3.86	3.17 - 4.56	3.36 - 4.36
BUN / UN3	mmol/l	6.10	5.19 - 7.02	5.46 - 6.74	15.4	13.1 - 17.7	13.8 - 17.0
	mg/dl	17.1	14.6 - 19.7	15.3 - 18.9	43.2	36.8 - 49.7	38.7 - 47.8
CALC / CA	mmol/l	3.00	2.73 - 3.27	2.82 - 3.18	3.34	3.04 - 3.64	3.14 - 3.54
	mg/dl	12.0	10.9 - 13.1	11.3 - 12.8	13.4	12.2 - 14.6	12.6 - 14.2
T-CHO / TC	mmol/l	3.22	2.90 - 3.54	2.99 - 3.45	4.25	3.83 - 4.68	3.95 - 4.55
	mg/dl	125	112 - 137	116 - 133	164	148 - 181	153 - 176
HDL-C / HDL	mmol/l	0.72	0.57 - 0.87		0.95	0.75 - 1.15	
	mg/dl	27.8	22.0 - 33.6		36.7	29.0 - 44.4	
CRE2	µmol/l	78.0	64.0 - 92.0	69.0 - 87.0	324	266 - 382	287 - 361
	mg/dl	0.88	0.72 - 1.04	0.78 - 0.98	3.67	3.01 - 4.32	3.24 - 4.09
GLUC	mmol/l	4.30	3.91 - 4.69	3.83 - 4.77	7.90	7.19 - 8.61	7.03 - 8.77
	mg/dl	77.5	70.5 - 84.5	69.0 - 86.0	142	130 - 155	127 - 158
MG	mmol/l	1.03	0.91 - 1.15	0.95 - 1.11	2.11	1.86 - 2.36	1.95 - 2.27
	mg/dl	2.50	2.20 - 2.80	2.32 - 2.69	5.13	4.51 - 5.74	4.74 - 5.51
PHOS / IP	mmol/l	0.90	0.77 - 1.04	0.82 - 0.98	1.70	1.45 - 1.96	1.55 - 1.85
	mg/dl	2.79	2.37 - 3.21	2.54 - 3.04	5.26	4.48 - 6.05	4.79 - 5.74
T-PRO / TP	g/l	56.0	49.3 - 62.7	52.6 - 59.4	67.0	59.0 - 75.0	63.0 - 71.0
	g/dl	5.60	4.93 - 6.27	5.26 - 5.94	6.70	5.90 - 7.50	6.30 - 7.10
TRIG / TG	mmol/l	0.83	0.65 - 1.01	0.76 - 0.90	2.05	1.68 - 2.42	1.87 - 2.23
	mg/dl	72.6	56.9 - 88	66.1 - 79.2	179	147 - 212	163 - 196
UA	µmol/l	340	299 - 381	316 - 364	451	397 - 505	419 - 483
	mg/dl	5.71	5.03 - 6.40	5.31 - 6.11	7.58	6.67 - 8.49	7.05 - 8.11

1 QUALAB Version 8.0, 1.9.2019

2 Deutsches Ärzteblatt 2014; Jg. 111, 19.9.2014

3 Labor und Diagnose, L. Thomas; BUN (Harnstoff-N) mg/dl x 2.14 = Harnstoff mg/dl

