



<b>LOT</b>		<b>MX 414 L</b>	<b>CONTROL</b>		<b>L</b>		 (Exp.)				<b>2019-02-05</b> (YYYY - MM - DD)	
		Rev 1	<b>LyseVet</b>									
PARAMETRES PARAMETERS	UNITES UNITS	MICROS									TOLERANCES TOLERANCE	
		Sci Vet ABC	Sci Vet ABC Plus ESV 60									
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>1.8</b>	<b>1.8</b>								± 0.4	
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>2.37</b>	<b>2.35</b>								± 0.15	
HB HGB	g/dl	<b>5.9</b>	<b>5.9</b>								± 0.4	
	g/l	<b>59</b>	<b>59</b>								± 4	
	mmol/l	<b>3.66</b>	<b>3.66</b>								± 0.25	
HT HCT	%	<b>16.6</b>	<b>16.2</b>								± 2.0	
	l/l	<b>0.166</b>	<b>0.162</b>								± 0.020	
VGM MCV	µm <sup>3</sup> ; fl	<b>70</b>	<b>69</b>								± 4	
TGMH MCH	pg	<b>24.9</b>	<b>25.1</b>								± 2.0	
	fmol	<b>1.55</b>	<b>1.56</b>								± 0.12	
CCMH MCHC	g/dl	<b>35.6</b>	<b>36.4</b>								± 3.0	
	g/l	<b>356</b>	<b>364</b>								± 30	
	mmol/l	<b>22.08</b>	<b>22.60</b>								± 1.86	
IDR RDW	%	N/A	<b>14.0</b>								± 3.0	
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>70</b>	<b>75</b>								± 20	
VPM MPV	µm <sup>3</sup> ; fl	N/A	<b>9.0</b>								± 2.0	
LYMPHO	#	N/A	<b>0.99</b>								± 0.40	
	%	N/A	<b>55.0</b>								± 8.0	
MONO	#	N/A	<b>0.21</b>								± 0.20	
	%	N/A	<b>11.5</b>								± 6.0	
GRANULO	#	N/A	<b>0.60</b>								± 0.40	
	%	N/A	<b>33.5</b>								± 7.0	

LOT		MX 414 N		CONTROL		N		(Exp.)		2019-02-05 (YYYY-MM-DD)	
Rev 1		LyseVet									
PARAMETRES PARAMETERS	UNITES UNITS	MICROS									TOLERANCES TOLERANCE
		Scil Vet ABC	Scil Vet ABC Plus ESV 60								
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	7.1	7.3								± 0.8
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	4.60	4.59								± 0.18
HB HGB	g/dl	13.2	13.2								± 0.5
	g/l	132	132								± 5
	mmol/l	8.20	8.20								± 0.31
HT HCT	%	36.8	36.7								± 2.5
	l/l	0.368	0.367								± 0.025
VGM MCV	µm <sup>3</sup> ; fl	80	80								± 4
TGMH MCH	pg	28.7	28.8								± 2.0
	fmol	1.78	1.79								± 0.12
CCMH MCHC	g/dl	35.9	35.9								± 3.0
	g/l	359	359								± 30
	mmol/l	22.28	22.32								± 1.86
IDR RDW	%	N/A	14.0								± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	245	252								± 40
VPM MPV	µm <sup>3</sup> ; fl	N/A	8.8								± 2.0
LYMPHO	#	N/A	2.12								± 0.40
	%	N/A	29.0								± 5.0
MONO	#	N/A	0.62								± 0.30
	%	N/A	8.5								± 4.0
GRANULO	#	N/A	4.56								± 0.60
	%	N/A	62.5								± 6.0

1300010912-A  
Ref: TEMP-0830 Rev.38

<b>LOT</b>		<b>MX 414 H</b>	<b>CONTROL H</b>							 (Exp.)	<b>2019-02-05</b> (YYYY - MM - DD)
		Rev 1	<b>LyseVet</b>								
PARAMETRES PARAMETERS	UNITES UNITS	MICROS									TOLERANCES TOLERANCE
		Sci Vet ABC	Sci Vet ABC Plus ESV 60								
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>19.6</b>	<b>20.2</b>								± 1.6
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>5.68</b>	<b>5.68</b>								± 0.20
HB HGB	g/dl	<b>18.0</b>	<b>18.0</b>								± 0.6
	g/l	<b>180</b>	<b>180</b>								± 6
	mmol/l	<b>11.18</b>	<b>11.18</b>								± 0.37
HT HCT	%	<b>50.6</b>	<b>50.6</b>								± 3.0
	l/l	<b>0.506</b>	<b>0.506</b>								± 0.030
VGM MCV	µm <sup>3</sup> ; fl	<b>89</b>	<b>89</b>								± 4
TGMH MCH	pg	<b>31.7</b>	<b>31.7</b>								± 2.0
	fmol	<b>1.97</b>	<b>1.97</b>								± 0.12
CCMH MCHC	g/dl	<b>35.6</b>	<b>35.6</b>								± 3.0
	g/l	<b>356</b>	<b>356</b>								± 30
	mmol/l	<b>22.11</b>	<b>22.11</b>								± 1.86
IDR RDW	%	N/A	<b>13.5</b>								± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>470</b>	<b>475</b>								± 45
VPM MPV	µm <sup>3</sup> ; fl	N/A	<b>8.6</b>								± 2.0
LYMPHO	#	N/A	<b>2.42</b>								± 1.00
	%	N/A	<b>12.0</b>								± 5.0
MONO	#	N/A	<b>1.01</b>								± 0.80
	%	N/A	<b>5.0</b>								± 4.0
GRANULO	#	N/A	<b>16.77</b>								± 1.40
	%	N/A	<b>83.0</b>								± 7.0