




<b>LOT</b>		<b>MX 418</b>	<b>L</b>	<b>CONTROL</b>		<b>L</b>	 <b>(Exp.)</b>				<b>2019-10-05</b> (YYYY - MM - DD)	
		Rev 1		<b>LyseVet</b>								
PARAMETRES PARAMETERS	UNITES UNITS	MICROS									TOLERANCES TOLERANCE	
		Sci Vet ABC	Sci Vet ABC Plus ESV 60									
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>1.8</b>	<b>1.8</b>								± 0.4	
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>2.34</b>	<b>2.29</b>								± 0.15	
HB HGB	g/dl	<b>5.9</b>	<b>5.9</b>								± 0.4	
	g/l	<b>59</b>	<b>59</b>								± 4	
	mmol/l	<b>3.66</b>	<b>3.66</b>								± 0.25	
HT HCT	%	<b>16.4</b>	<b>16.0</b>								± 2.0	
	l/l	<b>0.164</b>	<b>0.160</b>								± 0.020	
VGM MCV	µm <sup>3</sup> ; fl	<b>70</b>	<b>70</b>								± 4	
TGMH MCH	pg	<b>25.2</b>	<b>25.8</b>								± 2.0	
	fmol	<b>1.57</b>	<b>1.60</b>								± 0.12	
CCMH MCHC	g/dl	<b>36.0</b>	<b>36.8</b>								± 3.0	
	g/l	<b>360</b>	<b>368</b>								± 30	
	mmol/l	<b>22.37</b>	<b>22.86</b>								± 1.86	
IDR RDW	%	N/A	<b>14.0</b>								± 3.0	
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>68</b>	<b>75</b>								± 20	
VPM MPV	µm <sup>3</sup> ; fl	N/A	<b>8.4</b>								± 2.0	
LYMPHO	#	N/A	<b>0.99</b>								± 0.40	
	%	N/A	<b>55.0</b>								± 8.0	
MONO	#	N/A	<b>0.20</b>								± 0.20	
	%	N/A	<b>10.5</b>								± 6.0	
GRANULO	#	N/A	<b>0.62</b>								± 0.40	
	%	N/A	<b>34.5</b>								± 7.0	

<b>LOT</b>		<b>MX 418 N</b>	<b>CONTROL N</b>							 (Exp.)	<b>2019-10-05</b> (YYYY-MM-DD)
		Rev 1	<b>LyseVet</b>								
PARAMETRES PARAMETERS	UNITES UNITS	MICROS								TOLERANCES TOLERANCE	
		Scil Vet ABC	Scil Vet ABC Plus ESV 60								
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>7.2</b>	<b>7.3</b>							± 0.8	
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>4.62</b>	<b>4.60</b>							± 0.18	
HB HGB	g/dl	<b>13.2</b>	<b>13.2</b>							± 0.5	
	g/l	<b>132</b>	<b>132</b>							± 5	
	mmol/l	<b>8.20</b>	<b>8.20</b>							± 0.31	
HT HCT	%	<b>36.5</b>	<b>36.3</b>							± 2.5	
	l/l	<b>0.365</b>	<b>0.363</b>							± 0.025	
VGM MCV	µm <sup>3</sup> ; fl	<b>79</b>	<b>79</b>							± 4	
TGMH MCH	pg	<b>28.6</b>	<b>28.7</b>							± 2.0	
	fmol	<b>1.77</b>	<b>1.78</b>							± 0.12	
CCMH MCHC	g/dl	<b>36.2</b>	<b>36.3</b>							± 3.0	
	g/l	<b>362</b>	<b>363</b>							± 30	
	mmol/l	<b>22.46</b>	<b>22.56</b>							± 1.86	
IDR RDW	%	N/A	<b>14.0</b>							± 3.0	
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>238</b>	<b>247</b>							± 40	
VPM MPV	µm <sup>3</sup> ; fl	N/A	<b>8.4</b>							± 2.0	
LYMPHO	#	N/A	<b>2.12</b>							± 0.40	
	%	N/A	<b>29.0</b>							± 5.0	
MONO	#	N/A	<b>0.62</b>							± 0.30	
	%	N/A	<b>8.5</b>							± 4.0	
GRANULO	#	N/A	<b>4.56</b>							± 0.60	
	%	N/A	<b>62.5</b>							± 6.0	

Ref: TEMP-0830 Rev.39 BACK / VERSO 1300060626

<b>LOT</b>		<b>MX 418 H</b>	<b>CONTROL H</b>							 (Exp.)	<b>2019-10-05</b> (YYYY - MM - DD)
		Rev 1	<b>LyseVet</b>								
PARAMETRES PARAMETERS	UNITES UNITS	MICROS								TOLERANCES TOLERANCE	
		Scil Vet ABC	Scil Vet ABC Plus ESV 60								
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>19.9</b>	<b>20.2</b>							± 1.6	
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>5.60</b>	<b>5.62</b>							± 0.20	
HB HGB	g/dl	<b>17.9</b>	<b>17.9</b>							± 0.6	
	g/l	<b>179</b>	<b>179</b>							± 6	
	mmol/l	<b>11.12</b>	<b>11.12</b>							± 0.37	
HT HCT	%	<b>49.3</b>	<b>49.5</b>							± 3.0	
	l/l	<b>0.493</b>	<b>0.495</b>							± 0.030	
VGM MCV	µm <sup>3</sup> ; fl	<b>88</b>	<b>88</b>							± 4	
TGMH MCH	pg	<b>32.0</b>	<b>31.9</b>							± 2.0	
	fmol	<b>1.98</b>	<b>1.98</b>							± 0.12	
CCMH MCHC	g/dl	<b>36.3</b>	<b>36.2</b>							± 3.0	
	g/l	<b>363</b>	<b>362</b>							± 30	
	mmol/l	<b>22.56</b>	<b>22.48</b>							± 1.86	
IDR RDW	%	N/A	<b>13.5</b>							± 3.0	
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>505</b>	<b>522</b>							± 45	
VPM MPV	µm <sup>3</sup> ; fl	N/A	<b>8.3</b>							± 2.0	
LYMPHO	#	N/A	<b>2.53</b>							± 1.00	
	%	N/A	<b>12.5</b>							± 5.0	
MONO	#	N/A	<b>1.11</b>							± 0.80	
	%	N/A	<b>5.5</b>							± 4.0	
GRANULO	#	N/A	<b>16.56</b>							± 1.40	
	%	N/A	<b>82.0</b>							± 7.0	