




<b>LOT</b>		<b>MX 427 L</b>	<b>CONTROL</b>		<b>L</b>		 (Exp.)				<b>2021-04-05</b> (YYYY - MM - DD)	
		Rev 1	<b>LyseVet</b>									
PARAMETRES PARAMETERS	UNITES UNITS	MICROS									TOLERANCES TOLERANCE	
		Scil Vet ABC	Scil Vet ABC Plus ESV 60									
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	1.8	1.8								± 0.4	
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.35	2.35								± 0.15	
HB HGB	g/dl	5.9	5.9								± 0.4	
	g/l	59	59								± 4	
	mmol/l	3.66	3.66								± 0.25	
HT HCT	%	16.5	16.5								± 2.0	
	l/l	0.165	0.165								± 0.020	
VGM MCV	µm <sup>3</sup> ; fl	70	70								± 4	
TGMH MCH	pg	25.1	25.1								± 2.0	
	fmol	1.56	1.56								± 0.12	
CCMH MCHC	g/dl	35.9	35.9								± 3.0	
	g/l	359	359								± 30	
	mmol/l	22.27	22.27								± 1.86	
IDR RDW	%	N/A	15.0								± 3.0	
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	72	78								± 20	
VPM MPV	µm <sup>3</sup> ; fl	N/A	8.6								± 2.0	
LYMPHO	#	N/A	1.00								± 0.40	
	%	N/A	55.5								± 8.0	
MONO	#	N/A	0.20								± 0.20	
	%	N/A	11.0								± 6.0	
GRANULO	#	N/A	0.60								± 0.40	
	%	N/A	33.5								± 7.0	

<b>LOT</b>		<b>MX 427</b>	<b>N</b>	<b>CONTROL</b>		<b>N</b>					 (Exp.)	<b>2021-04-05</b> (YYYY-MM-DD)	
		Rev 1		<b>LyseVet</b>									
PARAMETRES PARAMETERS	UNITES UNITS	MICROS										TOLERANCES TOLERANCE	
		Scil Vet ABC	Scil Vet ABC Plus ESV 60										
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	7.1	7.2									± 0.8	
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	4.71	4.72									± 0.18	
HB HGB	g/dl	13.7	13.7									± 0.5	
	g/l	137	137									± 5	
	mmol/l	8.51	8.51									± 0.31	
HT HCT	%	38.2	38.2									± 2.5	
	l/l	0.382	0.382									± 0.025	
VGM MCV	µm <sup>3</sup> ; fl	81	81									± 4	
TGMH MCH	pg	29.1	29.0									± 2.0	
	fmol	1.81	1.80									± 0.12	
CCMH MCHC	g/dl	35.9	35.8									± 3.0	
	g/l	359	358									± 30	
	mmol/l	22.30	22.25									± 1.86	
IDR RDW	%	N/A	14.3									± 3.0	
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	260	280									± 40	
VPM MPV	µm <sup>3</sup> ; fl	N/A	8.5									± 2.0	
LYMPHO	#	N/A	1.98									± 0.40	
	%	N/A	27.5									± 5.0	
MONO	#	N/A	0.61									± 0.30	
	%	N/A	8.5									± 4.0	
GRANULO	#	N/A	4.61									± 0.60	
	%	N/A	64.0									± 6.0	

Ref: TEMP-0830 Rev.39 BACK / VERSO 1300060626

<b>LOT</b>		<b>MX 427 H</b>	<b>CONTROL H</b>							 (Exp.)	<b>2021-04-05</b> (YYYY-MM-DD)
		Rev 1	<b>LyseVet</b>								
PARAMETRES PARAMETERS	UNITES UNITS	MICROS								TOLERANCES TOLERANCE	
		Scil Vet ABC	Scil Vet ABC Plus ESV 60								
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>20.0</b>	<b>20.0</b>							± 1.6	
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>5.70</b>	<b>5.72</b>							± 0.20	
HB HGB	g/dl	<b>18.1</b>	<b>18.1</b>							± 0.6	
	g/l	<b>181</b>	<b>181</b>							± 6	
	mmol/l	<b>11.24</b>	<b>11.24</b>							± 0.37	
HT HCT	%	<b>50.7</b>	<b>50.9</b>							± 3.0	
	l/l	<b>0.507</b>	<b>0.509</b>							± 0.030	
VGM MCV	µm <sup>3</sup> ; fl	<b>89</b>	<b>89</b>							± 4	
TGMH MCH	pg	<b>31.8</b>	<b>31.6</b>							± 2.0	
	fmol	<b>1.97</b>	<b>1.97</b>							± 0.12	
CCMH MCHC	g/dl	<b>35.7</b>	<b>35.6</b>							± 3.0	
	g/l	<b>357</b>	<b>356</b>							± 30	
	mmol/l	<b>22.16</b>	<b>22.08</b>							± 1.86	
IDR RDW	%	N/A	<b>13.0</b>							± 3.0	
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>530</b>	<b>555</b>							± 45	
VPM MPV	µm <sup>3</sup> ; fl	N/A	<b>8.3</b>							± 2.0	
LYMPHO	#	N/A	<b>2.30</b>							± 1.00	
	%	N/A	<b>11.5</b>							± 5.0	
MONO	#	N/A	<b>1.00</b>							± 0.80	
	%	N/A	<b>5.0</b>							± 4.0	
GRANULO	#	N/A	<b>16.70</b>							± 1.40	
	%	N/A	<b>83.5</b>							± 7.0	