




<b>LOT</b>		<b>MX 428</b>	<b>L</b>	<b>CONTROL</b>		<b>L</b>					 (Exp.)	<b>2021-06-05</b> (YYYY-MM-DD)	
		Rev 1		<b>LyseVet</b>									
PARAMETRES PARAMETERS	UNITES UNITS	MICROS										TOLERANCES TOLERANCE	
		Scil Vet ABC	Scil Vet ABC Plus ESV 60										
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	1.7	1.8									± 0.4	
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.36	2.37									± 0.15	
HB HGB	g/dl	5.9	5.9									± 0.4	
	g/l	59	59									± 4	
	mmol/l	3.66	3.66									± 0.25	
HT HCT	%	16.5	16.4									± 2.0	
	l/l	0.165	0.164									± 0.020	
VGM MCV	µm <sup>3</sup> ; fl	70	69									± 4	
TGMH MCH	pg	25.0	24.9									± 2.0	
	fmol	1.55	1.55									± 0.12	
CCMH MCHC	g/dl	35.7	36.1									± 3.0	
	g/l	357	361									± 30	
	mmol/l	22.18	22.41									± 1.86	
IDR RDW	%	N/A	15.0									± 3.0	
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	61	78									± 20	
VPM MPV	µm <sup>3</sup> ; fl	N/A	8.7									± 2.0	
LYMPHO	#	N/A	1.05									± 0.40	
	%	N/A	58.5									± 8.0	
MONO	#	N/A	0.20									± 0.20	
	%	N/A	10.0									± 6.0	
GRANULO	#	N/A	0.57									± 0.40	
	%	N/A	31.5									± 7.0	

<b>LOT</b>		<b>MX 428 N</b>	<b>CONTROL N</b>							 (Exp.)	<b>2021-06-05</b> (YYYY-MM-DD)
		Rev 1	<b>LyseVet</b>								
PARAMETRES PARAMETERS	UNITES UNITS	MICROS								TOLERANCES TOLERANCE	
		Scil Vet ABC	Scil Vet ABC Plus ESV 60								
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>7.0</b>	<b>7.2</b>							± 0.8	
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>4.63</b>	<b>4.63</b>							± 0.18	
HB HGB	g/dl	<b>13.4</b>	<b>13.4</b>							± 0.5	
	g/l	<b>134</b>	<b>134</b>							± 5	
	mmol/l	<b>8.32</b>	<b>8.32</b>							± 0.31	
HT HCT	%	<b>37.5</b>	<b>37.5</b>							± 2.5	
	l/l	<b>0.375</b>	<b>0.375</b>							± 0.025	
VGM MCV	µm <sup>3</sup> ; fl	<b>81</b>	<b>81</b>							± 4	
TGMH MCH	pg	<b>28.9</b>	<b>28.9</b>							± 2.0	
	fmol	<b>1.80</b>	<b>1.80</b>							± 0.12	
CCMH MCHC	g/dl	<b>35.7</b>	<b>35.7</b>							± 3.0	
	g/l	<b>357</b>	<b>357</b>							± 30	
	mmol/l	<b>22.19</b>	<b>22.19</b>							± 1.86	
IDR RDW	%	N/A	<b>14.5</b>							± 3.0	
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>240</b>	<b>260</b>							± 40	
VPM MPV	µm <sup>3</sup> ; fl	N/A	<b>8.6</b>							± 2.0	
LYMPHO	#	N/A	<b>2.02</b>							± 0.40	
	%	N/A	<b>28.0</b>							± 5.0	
MONO	#	N/A	<b>0.58</b>							± 0.30	
	%	N/A	<b>8.0</b>							± 4.0	
GRANULO	#	N/A	<b>4.61</b>							± 0.60	
	%	N/A	<b>64.0</b>							± 6.0	

Ref: TEMP-0830 Rev.39 BACK / VERSO 1300060626

<b>LOT</b>		<b>MX 428 H</b>	<b>CONTROL H</b>							 (Exp.)	<b>2021-06-05</b> (YYYY-MM-DD)
		Rev 1	<b>LyseVet</b>								
PARAMETRES PARAMETERS	UNITES UNITS	MICROS									TOLERANCES TOLERANCE
		Scil Vet ABC	Scil Vet ABC Plus ESV 60								
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>19.3</b>	<b>19.7</b>								± 1.6
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>5.65</b>	<b>5.61</b>								± 0.20
HB HGB	g/dl	<b>18.0</b>	<b>18.0</b>								± 0.6
	g/l	<b>180</b>	<b>180</b>								± 6
	mmol/l	<b>11.18</b>	<b>11.18</b>								± 0.37
HT HCT	%	<b>50.3</b>	<b>49.9</b>								± 3.0
	l/l	<b>0.503</b>	<b>0.499</b>								± 0.030
VGM MCV	µm <sup>3</sup> ; fl	<b>89</b>	<b>89</b>								± 4
TGMH MCH	pg	<b>31.9</b>	<b>32.1</b>								± 2.0
	fmol	<b>1.98</b>	<b>1.99</b>								± 0.12
CCMH MCHC	g/dl	<b>35.8</b>	<b>36.1</b>								± 3.0
	g/l	<b>358</b>	<b>361</b>								± 30
	mmol/l	<b>22.23</b>	<b>22.39</b>								± 1.86
IDR RDW	%	N/A	<b>13.5</b>								± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>493</b>	<b>500</b>								± 45
VPM MPV	µm <sup>3</sup> ; fl	N/A	<b>8.7</b>								± 2.0
LYMPHO	#	N/A	<b>2.36</b>								± 1.00
	%	N/A	<b>12.0</b>								± 5.0
MONO	#	N/A	<b>0.99</b>								± 0.80
	%	N/A	<b>5.0</b>								± 4.0
GRANULO	#	N/A	<b>16.35</b>								± 1.40
	%	N/A	<b>83.0</b>								± 7.0