

LOT		MX 450 L		CONTROL		L		(Exp.)		2025-02-05	
		Rev 1								(YYYY-MM-DD)	
PARAMETRES PARAMETERS	UNITES UNITS	LyseVet								TOLERANCES TOLERANCE	
		MICROS									
			Scil Vet ABC Plus Scil Vet ABC ESV 60								
GB WBC	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l		1.8								± 0.4
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l		2.29								± 0.15
HB HGB	g/dl		5.9								± 0.4
	g/l		59								± 4
	mmol/l		3.66								± 0.25
HT HCT	%		16.0								± 2.0
	l/l		0.160								± 0.020
VGM MCV	µm <sup>3</sup> ; fl		70								± 4
TGMH MCH	pg		25.8								± 2.0
	fmol		1.60								± 0.12
CCMH MCHC	g/dl		36.8								± 3.0
	g/l		368								± 30
	mmol/l		22.86								± 1.86
IDR RDW *	%		14.5								± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l		73								± 20
VPM MPV *	µm <sup>3</sup> ; fl		8.0								± 2.0
LYMPHO *	#		1.03								± 0.40
	%		57.0								± 8.0
MONO *	#		0.20								± 0.18
	%		11.0								± 9.9
GRANULO *	#		0.58								± 0.40
	%		32.0								± 10.0

\* N/A for Micros Scil Vet ABC

**HORIBA ABX SAS**  
HORIBA ABX SAS  
Parc Euromédecine  
Rue du Caducée - BP 7290  
34184 MONTPELLIER Cedex 4  
FRANCE  
Tél.: 33 (0) 4 67 14 15 16  
Fax.: 33 (0) 4 67 14 15 17

<b>LOT</b>		<b>MX 450 N</b>	<b>CONTROL N</b>							(Exp.)	<b>2025-02-05</b> (YYYY-MM-DD)
		Rev 1	<b>LyseVet</b>								
PARAMETRES PARAMETERS	UNITES UNITS	MICROS									TOLERANCES TOLERANCE
			Scil Vet ABC Plus Scil Vet ABC ESV 60								
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l		<b>7.0</b>								± 0.8
GR RBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l		<b>4.58</b>								± 0.18
HB HGB	g/dl		<b>13.2</b>								± 0.5
	g/l		<b>132</b>								± 5
	mmol/l		<b>8.20</b>								± 0.31
HT HCT	%		<b>36.6</b>								± 2.5
	l/l		<b>0.366</b>								± 0.025
VGM MCV	µm <sup>3</sup> ; fl		<b>80</b>								± 4
TGMH MCH	pg		<b>28.8</b>								± 2.0
	fmol		<b>1.79</b>								± 0.12
CCMH MCHC	g/dl		<b>36.0</b>								± 3.0
	g/l		<b>360</b>								± 30
	mmol/l		<b>22.37</b>								± 1.86
IDR RDW *	%		<b>14.0</b>								± 3.0
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l		<b>260</b>								± 40
VPM MPV *	µm <sup>3</sup> ; fl		<b>7.9</b>								± 2.0
LYMPHO *	#		<b>2.03</b>								± 0.40
	%		<b>29.0</b>								± 5.0
MONO *	#		<b>0.53</b>								± 0.48
	%		<b>7.5</b>								± 6.8
GRANULO *	#		<b>4.45</b>								± 0.70
	%		<b>63.5</b>								± 10.0

\* N/A for Micros Scil Vet ABC

1300104143 BACK / VERSO Ref: TEMP-0630 Rev.43

LOT

MX 450 H  
Rev 1

CONTROL

H

 (Exp.)

2025-02-05  
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	LyseVet								TOLERANCES TOLERANCE
		MICROS								
			Scil Vet ABC Plus Scil Vet ABC ESV 60							
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l		18.9							± 1.6
GR RBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l		5.72							± 0.20
HB HGB	g/dl		18.0							± 0.6
	g/l		180							± 6
	mmol/l		11.18							± 0.37
HT HCT	%		50.9							± 3.0
	l/l		0.509							± 0.030
VGM MCV	µm <sup>3</sup> ; fl		89							± 4
TGMH MCH	pg		31.5							± 2.0
	fmol		1.95							± 0.12
CCMH MCHC	g/dl		35.4							± 3.0
	g/l		354							± 30
	mmol/l		21.96							± 1.86
IDR RDW *	%		13.5							± 3.0
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l		514							± 45
VPM MPV *	µm <sup>3</sup> ; fl		7.7							± 2.0
LYMPHO *	#		2.55							± 1.00
	%		13.5							± 5.0
MONO *	#		0.76							± 0.68
	%		4.0							± 3.6
GRANULO *	#		15.59							± 1.89
	%		82.5							± 10.0

\* N/A for Micros Scil Vet ABC